

DST-PURSE PROJECT - MAHATMA GANDHI UNIVERSITY

SCHOOL OF BIOSCIENCES

KOTTAYAM-686560

TENDER NOTICE FOR PURCHASE OF EQUIPMENT

TENDER NOTICE No: MGU/DST-PURSE/SM/2018/01

Date: 10.01.2018

Competitive tender in sealed cover for supply of **Stereomicroscope** is invited from reputed firms and authorized dealers so as to reach the office of the under signed not later than **3 p.m. on 24.01.2018**. Technical bid of the tenders will be opened on **25.01.2018** at **3.00 p.m.** in the presence of authorized representatives of tendered firms. The financial bid of the Tender will be opened on **25.01.2018** at 3pm

| Sl. No. | Scheduled Item | Specifications | E.M.D Rs | Tender Fees |
|---------|------------------|--|------------------------|-------------|
| 01 | Stereomicroscope | Annexure IV for detailed Specification | 2.5 % of the Estimated | 2250 |
| | | | | |

Co-ordinator
DST-Purse Project
School of Biosciences,
M. G University Kottayam
Kerala , India

NOTE: (1) PLEASE SEE THE ANNEXURES I, ANNEXURE II, ANNEXURE III and

ANNEXURE IV ATTACHED BELOW

(2) ALL THE DOCUMENTS RELATED TO THE TENDER CAN BE DOWNLOADED FROM THE UNIVERSITY WEBSITE- www.mgu.ac.in, in the section "Quotation/tender"

ANNEXURE I

General Terms and Conditions

1. The main envelop should be super scribed: **“Tender for Equipment procurement under DST-PURSE Project ”**
2. The non refundable application fee of accompanied with tender for each item. The application fee must b e as **D.D. drawn in favour of Finance offi cer, M.G University.**
3. Tenders must accompany a copy of the “General Terms and Condit ions, Annexure I and Annexure II and III” section of this document, signed and stamped on each page indicating that they agree to these.
4. Last date of submission of tender along with requisite fee, EMD and all documents is **24.01.2018**. The DD f or EMD or any other accepted document must be drawn in favor of

Co-ordinator, DST-PURSE project, M. G University, Kottayam, Kerala payable at SBT Mahatma Gandhi University Campus Branch.

5. All other charges including VAT/CST, Excise Duty and other levies payable by C.I.P should be clearly indicated otherwise it will be presumed that the rates quoted are inclusive of all these charges and will not be paid.
6. The Excise Duty component (with percentage) should be indicated , as the University is exempted from the payment of Custom/Excise Duty. Exemption will be availed by providing Custom/Excise Duty Exemption Certificate with order.
7. The Delivery Schedule, Payment Terms & Warranty/Guarantee etc must be clearly indicated in the technical bid. The charges for extended warranty and/or Annual Maintenance Contract after the expiry of offered warranty period should also be specified in the financial bid.
8. The manufacturers’ printed literature/catalogue/drawing/user’s list in respect of range of product being quoted should also be submitted with the offer.
9. Our Institute is registered with the Department of Scientific & Industrial Research (DSIR), so Excise duty exemption will be provided.

Other requirements for delivery and complete installation

1. Delivery at School o f Environmental Sciences, Mahatma Gandhi University, Kottayam
2. All other requirements for satisfactory installation of **equipment**.
3. It will be the responsibility of the supplier to deliver the ordered materials at School of Environmental Sciences, Mahatma Gandhi University, Kottayam.
4. All required materials for satisfactory installation are to be provided by the supplier at their own cost.

ANNEXURE II:
TENDER FORM PART-I (TECHNICAL BID)

PART-I (TECHNICAL BID) OF TENDER NO:

Last date for receipt:

Due date for opening Part –I (TECHNICAL BID):

Tenderer's Offer No:

Date:

From,

M/s.
.....

To,

Dr. Keerthi TR,

Co-ordinator

DST-Purse Project

School of Biosciences, M. G. University,

Kottayam, Kerala, India

Dear Sir,

I/We have gone through the tendering conditions pertaining to the Tender and General Terms and Conditions of Contract and other requirement for delivery and complete Installation and Special Conditions of Contract contained herein with this tender document. I/we hereby agree to supply the stores conforming to the tender specifications incorporated in ANNEXURE I of the tender document and also agree to abide by your General Conditions of all Contracts and Special Conditions of Contract contained in the ANNEXURE I of the Tender document.

You will be at liberty to accept any or more of the items of stores offered by us and I/we shall be bound to supply you the stores as may be specified in the Purchase Order/Contract.

I/We hereby agree to keep the price valid for your acceptance for a period of 30 days from the date of opening of Part-II (Financial bid) of the tender

I/We are also enclosing herewith all the leaflets catalogue etc. pertaining to the stores offered.

Yours faithfully

Stamp and Signature of the Tenderer

ANNEXURE III: TENDER FORM PART-II (FINANCIAL BID)

PART-II (FINANCIAL BID) OF TENDER NO:

Last date for receipt:

Due date for opening Part –II (FINANCIAL BID):

Tenderer’s Offer No:

Date:

From,

M/s.
.....
.....

To,

Dr. Keerthi TR,
Co-ordinator
DST-Purse Project
School of Biosciences, M. G. University,
Kottayam, Kerala, India

Dear Sir,

In response to your invitation and as per your tendering and contracting conditions, the prices applicable for the scope of supply contained in ANNEXURE-I (TECHNICAL BID) of our tender are indicated in the format at annexure “A” to this tender.

We hereby agree to keep the price valid for your acceptance for a period of 30 days from the date of actual opening of Part-II (FINANCIAL BID) of the tender.

Yours faithfully,

Stamp and Signature of the Tenderer

ANNEXURE IV

Detailed Specifications

| Name of the equipment | Specification |
|-------------------------|---|
| Stereomicroscope | <ol style="list-style-type: none"> 1) Trinocular, imported stereo microscope body with inbuilt objective , apochromatic corrected to get sharp image 2) Total Magnification: 40x (or) more 3) The eyepieces should be 10x/22 Br.foc (or) more Br. foc 4) Zoom ratio: 1 : 5 to 1 : 8 5) Zoom adjustable knob : left/right , single shaft horizontal knob, 6) Working distance: 90mm or more 7) Light source: inbuilt LED source with double spot and ring light illuminator 8) Microscope stand should have transmitted and reflected light illumination, upgradable to oblique illumination. 9) Dedicated CMOS camera with 5 MP or more, should have option for video recording , with suitable advanced image processing software. 10) All camera specifications such as: pixel count and resolution, sensor size, exposure etc. should be clearly mentioned in the quote. 11) Microscopic camera and microscope should be from same manufacturer 12) Imported , should meet all international and safety standards 13) A suitable table top computer for advanced image processing/ analysis with appropriate software may |

| | |
|--|---|
| | <p>be quoted as optional accessory. Minimum specification for computer should be Branded - i3,4gb,1tb,DVD writer, Cabinet with SMPS with antivirus software, 22' monitor.</p> |
|--|---|